Bacteriological Water Potability Test For E. coli

To take a sample:

- 1. You must obtain sterile collection containers (sterile bottles or bags) from Roesink Laboratories, LLC or purchase your own from a supplier.
- 2. Do not rinse container
- 3. Obtain sample from a stationary faucet without an aerator or screen. Do not use a swivel faucet. Use a bleach solution to clean the opening where water leaves the faucet.
- 4. Run cold water at least five minutes before taking a sample.
- 5. Fill collection bottle past the 100ml mark but not to the top and replace lid at once. Do not touch the inside of the lid or bottle. The collection bottle contains a sodium thiosulfate tablet. DO NOT REMOVE THE TABLET.
- 6. Sample must be received by our lab within 24 hours of sampling. Refrigerate sealed bottle until delivery to the lab.

To obtain results:

Results are available as early as 48 hours after the sample is received by the lab. All results will be mailed or E-mailed, and you may call for results. If you have any questions, or for results, please call (330) 526-1914.